

Rely+On® PeraSafe®

High Level Disinfectant for
use with Medical Devices



- Outstanding biocidal action
- Optimised corrosion inhibitor system
- Suitable for the manual disinfection of endoscopes, ultra-sound probes and other reusable, heat sensitive medical devices
- Good environmental and operator safety profiles
- Easy to use unit dose sachet available

Rely+On® PeraSafe®

Specifically developed to offer rapid broad spectrum decontamination against viruses, mycobacteria, bacteria, yeasts and bacterial spores including *C. difficile*. Rely+On® PeraSafe®, with its optimised corrosion inhibitor system and near neutral pH, is a suitable solution for endoscope, ultra-sound probe and other reusable, heat sensitive medical device disinfection.

Biocidal Efficacy

Rely+On® PeraSafe® powder generates peracetic acid *in-situ* when dissolved in water. The neutral pH solution combines the biocidal efficacy of peracetic acid with excellent materials compatibility to allow repeated, rapid decontamination of pathogenic micro-organisms from heat labile, reusable medical devices and instruments.

Proven Broad Spectrum Efficacy

Independently proven & highly effective, Rely+On® PeraSafe® provides a rapid broad spectrum decontamination against medical disease-causing pathogens, such as Hepatitis B & C, Mycobacteria (including tuberculosis), *Pseudomonas aeruginosa*, *Clostridium difficile* (spores), MRSA and *Candida albicans*.



Rely+On® PeraSafe® has been independently evaluated by various global reference testing laboratories, with full proven efficacy in 10 minutes* against these key medical pathogen types. These results clearly demonstrate Rely+On® PeraSafe®'s rapid biocidal action and suitability for use in medical device disinfection.

Medical Device Compatibility

Rely+On® PeraSafe® is compatible with rigid and flexible endoscopes, catheters, transducers and other thermolabile instruments and equipment.*

Certification

Rely+On® PeraSafe® complies with the Medical Device Regulation* as assessed by BSI Netherlands (Notified Body No. 2797) (*MDR Certificate 781707*).

* Contact Antec International Ltd for further details.

Operator Safety

Rely+On® PeraSafe® is not classified as a sensitiser in both the powdered form and the 16.2g per litre in-use dilution.

Environmental Profile

Rely+On® PeraSafe® has been formulated to include ingredients that have been carefully selected for their ability to degrade naturally within the environment.

Lifespan of Disinfectant Solution

Solutions of Rely+On® PeraSafe® can be used for up to 20 immersions or a maximum of 12 hours (when stored between 15-25°C). Unused or inactive solutions may be disposed of via the normal waste water discharge channels*.

Rely+On® PeraSafe® is available in the following presentations:

Pack Size	Volume of disinfectant solution
16.2 gram (sachet, unit dose)	1 litre
81 gram (bottle, unit dose)	5 litres
162 gram (bottle)	10 litres
810 gram (bottle)	50 litres

* Always check and observe the local site requirements concerning the disposal of solutions entering waste water treatment facilities. Operatives should always further check that the quantity of powder to be added is appropriate to the volume of water used – always refer to the instructions in the product instruction for use leaflet (IFU).



Easy to Prepare




Step 1

Fill a container with the appropriate volume of lukewarm tap water (30-35°C).



Step 2

Prepare the solution by adding 16.2 grams of Rely+On® PeraSafe® powder to every 1 litre of water.



Step 3

Stir until the powder dissolves to form a blue solution. Then leave for 15 minutes to allow the disinfectant solution to fully activate.

The blue color fades after preparation, and is not considered as a measure of solution validity.

The tables below summarise independent efficacy data of Rely+On® PeraSafe® against both reference test organisms, and the important disease-causing organisms of concern for disinfection of medical instruments.

Virucidal Efficacy

Organism	Strain	Test Protocol	Contact Time	Log Reduction
Adenovirus	Type 5	EN 14476	10 minutes	>4 log
Hepatitis B	DHBV strain	EN 14476	10 minutes	>4 log
Hepatitis C	BVD surrogate strain	EN 14476	10 minutes	>4 log
Poliovirus	Type 1	EN 14476	10 minutes	>4 log
Murine norovirus	S99 Berlin	EN 14476	10 minutes	>4 log
Adenovirus	Type 5	EN 17111	10 minutes	>4 log
Murine norovirus	S99 Berlin	EN 17111	10 minutes	>4 log
HIV	HTLV-III B	USA EPA 40 part 158	5 minutes	>4 log

Mycobactericidal Efficacy

Organism	Strain	Test Protocol	Contact Time	Log Reduction
Mycobacterium avium	ATCC 15769	EN 14348	10 minutes	>4 log
Mycobacterium terrae	ATCC 15755	EN 14348	10 minutes	>4 log
Mycobacterium avium	ATCC 15769	EN 14563	10 minutes	>4 log
Mycobacterium terrae	ATCC 15755	EN 14563	10 minutes	>4 log

Bactericidal Efficacy

Organism	Strain	Test Protocol	Contact Time	Log Reduction
Acinetobacter baumannii	Drug resistant DSMZ 102930	EN 13727	1 minute	>5 log
Campylobacter jejuni	Wild strain	EN 1276	5 minutes	>5 log
Escherichia coli	NCTC 12900	EN 1276	5 minutes	>5 log
Klebsiella aerogenes	NCTC 9528	EN 1276	5 minutes	>5 log
Staphylococcus aureus subsp. aureus	Drug resistant ATCC 335491 (MRSA)	EN 13727	1 minute	>5 log
Staphylococcus aureus	MRSA NCTC 12493	EN 1276	5 minutes	>5 log
Staphylococcus aureus	EMRSA type 15	EN 1276	5 minutes	>5 log
Salmonella typhimurium	Phage type 104	EN 1276	5 minutes	>5 log
Enterococcus hirae	NCTC 12367	EN 13727	10 minutes	>5 log
Staphylococcus aureus	NCTC 10788	EN 13727	10 minutes	>5 log
Pseudomonas aeruginosa	NCTC 6749	EN 13727	10 minutes	>5 log
Enterococcus hirae	ATCC 10541	EN 14561	10 minutes	>5 log
Staphylococcus aureus	ATCC 6538	EN 14561	10 minutes	>5 log
Pseudomonas aeruginosa	ATCC 15442	EN 14561	10 minutes	>5 log
Pseudomonas aeruginosa	NCTC 6749	Endoscope in-use trial	4 minutes	>7 log

Bactericidal Spore Efficacy

Organism	Strain	Test Protocol	Contact Time	Log Reduction
Bacillus subtilis var niger (spores)	ATCC 6633	EN 17126	10 minutes	>4 log
Clostridium difficile	NCTC 13366	EN 17126	10 minutes	>4 log

Yeasticidal Efficacy

Organism/Disease	Strain	Test Protocol	Contact Time	Log Reduction
Candida albicans	ATCC 10231	EN 13624	10 minutes	>4 log
Candida albicans	ATCC 10231	EN 14562	10 minutes	>4 log



Antec International Limited
 Windham Road, Chilton Industrial Estate
 Sudbury, Suffolk CO10 2XD, United Kingdom
 Tel: +44(0)1787 377305
 relyondisinfection@lanxess.com
 www.relyondisinfection.com

©2025 LANXESS. LANXESS and the LANXESS Logo, Rely+On® PeraSafe® are trademarks of LANXESS Deutschland GmbH or its affiliates. All trademarks are registered in many countries worldwide.

This information and our technical advice – whether verbal, in writing or by way of trials – is subject to change without notice and given in good faith but without warranty or guarantee, express or implied, and this also applies where proprietary rights of third parties are involved. Our advice does not release you from the obligation to verify the information currently provided - especially that contained in our safety data and technical information sheets - and to test our products as to their suitability for the intended processes and uses. The application, use and processing of our products and the products manufactured by you on the basis of our technical advice are beyond our control and, therefore, entirely your own responsibility. Our products are sold in accordance with the current version of our General Conditions of Sale and Delivery. The specified uses and registered claims for the product may vary from country to country. Please contact LANXESS to verify country-specific approved uses.